



Certificate of Analysis

1725 Brannum Lane
Yellow Springs, Ohio
45387-1107 USA
+1 937-767-7241

Product Category	ISE Reagent
Product Model Number	2972 (500 mg/L Ammonium, 10 mg/L Potassium)
Lot Number	25F100233
Method	Comparative ISE testing of Production reagent vs. Metrology reagent

Testing Results

Chemistry Tested	Determination	Tolerance	Result
Ammonium	(Average of Production) – (Average of Metrology)	$\leq 10 \text{ mg/L (2\%)}$	0.19
	Coefficient of Variation (CV) for Production reagent	$\leq 3\%$	0.26

The procedures used to manufacture and test this product have been verified by QA for GMP compliance.

Technician

Amanda Fenner

Date

6/11/2025