

Certificate of Analysis

**YSI Incorporated**

Yellow Springs Instrument Co., Yellow Springs, Ohio 45387 USA

Phone 937 767-7241 • 800 765-4974 • Fax 937 767-9353

a **xylem** brand

Product Category	Membrane
Product Model Number	2702
Lot Number	25F100466
Method	Amperometric testing with standards

Time of Testing	Determination	Specification	Result
Initial	Average calibration sensitivity	10 - 30 nA	10.79 nA
	Average Linearity	23.75 – 26.25 g/L	25.07 g/L
	Potassium Ferrocyanide interference	N/A	N/A
Day = _____	Average calibration sensitivity	N/A	N/A
	Average Linearity	N/A	N/A
	Potassium Ferrocyanide interference	N/A	N/A

The procedures used to manufacture and test this product have been verified by QA for compliance. The electronic signature below is representative of the Chemistry Technician testing and verifying conformity to acceptance criteria.

Completed by: Christy Woodford Date: 6/26/2025

