

# Certificate of Analysis



**YSI Incorporated**

**Yellow Springs Instrument Co., Yellow Springs, Ohio 45387 USA**

**Phone 937 767-7241•800 765-4974•Fax 937 767-9353**

Product Category	Bioanalytical Standard
Product Model Number	2367
Lot Number	25G100021
Method	Comparative enzymatic testing of Production standard vs Metrology standard

Chemistry Tested	Determination	Tolerance	Result
GLUCOSE	(Average of Production)-(Average of Metrology)	$\pm 0.05\text{mmol/L}$	0.03mmol/L
	Coefficient of variation for production standard (CV,%)	$\leq 2.0\%$	0.9 %

*The procedures used to manufacture and test this product have been verified by QA for compliance. The electronic signature below is representative of the Chemistry Technician testing and verifying conformity to acceptance criteria.*

Completed by: Amanda Fenner

Date: 7/3/2025

