

Certificate of Analysis



Yellow Springs Instrument Co., Yellow Springs, Ohio 45387 USA

Phone 937 767-7241 • 800 765-4974 • Fax 937 767-9353

Product Category	Sucrose Membrane
Product Model Number	2703
Lot Number	25H100154
Method	Amperometric testing with standards

Time of Testing	Determination	Specification	Results
Initial	Average Calibration Sensitivity	8.0-40.0 nA	9.7 nA
	Average Linearity	23.75-26.25 g/L	25.28 g/L
	Potassium Ferrocyanide Interference	0.0-0.3	0.00 g/L

The procedures used to manufacture and test this product have been verified by QA for compliance. The electronic signature below is representative of the Chemistry Technician testing and verifying conformity to acceptance criteria.

Completed by: Christy Woodford Date: 8/14/2025
Your Name _____

Date: 8/14/2025

