

# Certificate of Analysis



**YSI Incorporated**

Yellow Springs Instrument Co., Yellow Springs, Ohio 45387 USA

Phone 937 767-7241 • 800 765-4974 • Fax 937 767-9353

---

Product Category	Membrane
Product Model Number	2701
Lot Number	26C100049
Method	Amperometric testing with standards

---

Time of Testing	Determination	Specification	Results
Initial	Average Calibration Sensitivity	10 - 50 nA	15.35 nA
	Average Linearity	N/A	N/A
	Potassium Ferrocyanide Interference	-1.00 - 1.00	0.12 nA

---

*The procedures used to manufacture and test this product have been verified by QA for compliance. The electronic signature below is representative of the Chemistry Technician testing and verifying conformity to acceptance criteria.*

Completed by: Christy Woodford  
Your Name

Date: 3/11/2026

