

Certificate of Analysis



YSI Incorporated

Yellow Springs Instrument Co., Yellow Springs, Ohio 45387 USA

Phone 937 767-7241 • 800 765-4974 • Fax 937 767-9353

Product Category Membrane
Product Model Number 2725
Lot Number 26D100216
Method Amperometric testing with standards

| Time of Testing | Determination | Specification | Results |
|-----------------|---------------------------------|-----------------|-----------|
| Initial | Average Calibration Sensitivity | 5.0-25.0 nA | 6.71 nA |
| | Average Linearity | 2.375-2.625 g/L | 2.537 g/L |
| | Potassium Ferrocyanide | -0.2-0.2 g/L | 0.00 g/L |
| Final | Average Calibration Sensitivity | 5.0-25.0 nA | 6.50 g/L |
| | Average Linearity | 2.375-2.625 g/L | 2.517 g/L |
| | Potassium Ferrocyanide | -0.2-0.2 g/L | 0.00 g/L |

The procedures used to manufacture and test this product have been verified by QA for compliance. The electronic signature below is representative of the Chemistry Technician testing and verifying conformity to acceptance criteria.

Completed by: Christy Woodford
Your Name

Date: 4/20/2026

