

Certificate of Analysis



YSI Incorporated

Yellow Springs Instrument Co., Yellow Springs, Ohio 45387 USA

Phone 937 767-7241 • 800 765-4974 • Fax 937 767-9353

Product Category Buffer and solid products
Product Model Number 2392C
Lot Number 26E100008
Method Verification of mass

Determination	Specification	Result
Average Mass	0.20-0.30 grams	0.25 g

The procedures used to manufacture and test this product have been verified by QA for compliance. The electronic signature below is representative of the Chemistry Technician testing and verifying conformity to acceptance criteria.

Completed by:

Megan Durkee

A handwritten signature in black ink, appearing to read 'Megan Durkee', written over a horizontal line.

Date: 5/11/2026

