Certificate of Analysis

YSI Incorporated
Yellow Springs Instrument Co., Yellow Springs, Ohio 45387 USA
Phone 937 767-7241•800 765-4974•Fax 937 767-9353

a xylem brand

Product Category Membrane

Product Model Number 2761

Lot Number 26E100143

Method Amperometric testing with standards

Time of Testing	Determination	Specification	Result
Initial	Average calibration sensitivity	10 - 60 nA	63.05 nA
	Average Linearity	27.0 – 33.0 g/L	28.5 g/L
	Potassium Ferrocyanide interference	0 – 0.07 g/L nA	0.00 nA
Day = _____	Average calibration sensitivity	N/A	N/A
	Average Linearity	N/A	N/A
	Potassium Ferrocyanide interference	N/A	N/A

The procedures used to manufacture and test this product have been verified by QA for compliance. The electronic signature below is representative of the Chemistry Technician testing and verifying conformity to acceptance criteria.

Completed by: Christy Woodford Date: 5/14/2026

ISO 9001 CERTIFIED YSI INCORPORATED