



155 Hidden Ravines Dr
Powell, OH 43065
USA

Certificate of Analysis

Tel: 800.858.9682 . 740.881.5501

Fax: 740.881.5989

www.gfschemicals.com

Product name: YSI 6074 Turbidity Standard

Item #: 8174

Lot #: 24029392

Certified Values:

Specifications	Status	Results
Turbidity on YSI EXO Sensor (999.90-1020.1 FNU)	Pass	1008.89 FNU
Turbidity on YSI 6136 Sensor (905.98-962.02 NTU)	Pass	943.4 NTU
Absorbance at 455nm with 10mm pathlength	Report	1.0965

Comments YSI Equivalent Lot: 26C24029392

CoA #: COA-123628

Expiration date: March 2027

Certificate Created By: Sasikala Sellathurai

Print Date: March 10, 2026

Certified by: Aron Becza - Quality Assurance
Manager

Aron Becza

Traceability:

This material was processed under a quality management system that is registered to ISO 9001 2015. The equipment used in the testing of materials is NIST traceable. In cases where NIST traceability is not possible equipment manufactured recommendations and /or industry best practices are followed.

The following tests are performed under an A2LA accredited ISO/IEC 17025 management system:

- UV/VIS Spectrophotometer testing is performed according to ASTM method E169-16
- Turbidity according to ASTM methods D6855-17 and ASTM D7315-17

Samples for testing are obtained using GFS procedure GFSW-LAB-PPG-0022.

Testing is performed in a laboratory temperature of 22 °C +/- 2 °C.

Turbidity Instrument – A formazin calibrated turbidity meter was used in the formulation and testing of the standards for each specific instrument design.

Certified absorbance value is only valid when AMCO Clear 0.0 NTU is used for the instrument baseline.

UV/VIS Spectrophotometer Information – Certified absorbance readings obtained on either a Shimadzu UV-2600 spectrophotometer or a Shimadzu UV-2700 spectrophotometer that is verified yearly with NIST SMRs 2031, 2034 and Hg Lamp.

Measurement Uncertainty:

The uncertainty of the certified values is stated below. These values are to be input into the equation that follows to determine the acceptable tolerance:

- If absorbance is $0.0 < x < 0.5$ AU: then uncertainty is +/- 0.005 AU
- If absorbance is $0.5 < x < 1.5$ AU: then uncertainty is +/- 0.008 AU
- If absorbance is $1.5 < x < 4$ AU: then uncertainty is +/- 0.012 AU

Tolerance Statement

The difference between the mean measured turbidity value and the verified value provided on this certificate should not exceed the tolerance value.

Specifications:

Hazardous Information: Consult SDS for information on this product.

Storage Information: This standard should be stored at room temperature before and after opening. At no time should it be frozen or stored in extreme heat. These conditions can cause instability of the materials and lead to incorrect results.

Usage Information: This material is intended for use as a calibration or validation standard. For primary standards, pour necessary amounts of material into a sample container and run the sample per instrument manufacturer's user manual. Dispose of used nonhazardous materials. Do not pour back into bottle. These materials are not for dilution and do not require inversion or mixing to use unless otherwise stated on the bottle. Do not shake the standards as this can introduce greater error in the results. For secondary standards, place the vial in the instrument and run the sample per the manufacturer's user manual.

Not for direct use in food, cosmetics, finished pharmaceuticals or drug products. Supplier is not responsible for compliance with FDA Current Good Manufacturing Practices (cGMP), including without limitation for those finished drug products in 21 CFR Parts 210 and 211. Consult warranty limitations at www.gfschemicals.com

24029392



155 Hidden Ravines Dr
Powell, OH 43065
USA

Certificate of Analysis

Tel: 800.858.9682 . 740.881.5501

Fax: 740.881.5989

www.gfschemicals.com

Product name: YSI 6074 Turbidity Standard

Item #: 8174

Lot #: 24029602

Certified Values:

Specifications	Status	Results
Turbidity on YSI EXO Sensor (999.90-1020.1 FNU)	Pass	1013.38 FNU
Turbidity on YSI 6136 Sensor (905.98-962.02 NTU)	Pass	941.7 NTU
Absorbance at 455nm with 10mm pathlength	Report	1.0954

Comments YSI Equivalent Lot: 26C24029602

CoA #: COA-123827

Expiration date: March 2027

Certificate Created By: Sasikala Sellathurai

Print Date: March 16, 2026

Certified by: Aron Becza - Quality Assurance
Manager

Aron Becza

Traceability:

This material was processed under a quality management system that is registered to ISO 9001 2015. The equipment used in the testing of materials is NIST traceable. In cases where NIST traceability is not possible equipment manufactured recommendations and /or industry best practices are followed.

The following tests are performed under an A2LA accredited ISO/IEC 17025 management system:

- UV/VIS Spectrophotometer testing is performed according to ASTM method E169-16
- Turbidity according to ASTM methods D6855-17 and ASTM D7315-17

Samples for testing are obtained using GFS procedure GFSW-LAB-PPG-0022.

Testing is performed in a laboratory temperature of 22 °C +/- 2 °C.

Turbidity Instrument – A formazin calibrated turbidity meter was used in the formulation and testing of the standards for each specific instrument design.

Certified absorbance value is only valid when AMCO Clear 0.0 NTU is used for the instrument baseline.

UV/VIS Spectrophotometer Information – Certified absorbance readings obtained on either a Shimadzu UV-2600 spectrophotometer or a Shimadzu UV-2700 spectrophotometer that is verified yearly with NIST SMRs 2031, 2034 and Hg Lamp.

Measurement Uncertainty:

The uncertainty of the certified values is stated below. These values are to be input into the equation that follows to determine the acceptable tolerance:

- If absorbance is $0.0 < x < 0.5$ AU: then uncertainty is +/- 0.005 AU
- If absorbance is $0.5 < x < 1.5$ AU: then uncertainty is +/- 0.008 AU
- If absorbance is $1.5 < x < 4$ AU: then uncertainty is +/- 0.012 AU

Tolerance Statement

The difference between the mean measured turbidity value and the verified value provided on this certificate should not exceed the tolerance value.

Specifications:

Hazardous Information: Consult SDS for information on this product.

Storage Information: This standard should be stored at room temperature before and after opening. At no time should it be frozen or stored in extreme heat. These conditions can cause instability of the materials and lead to incorrect results.

Usage Information: This material is intended for use as a calibration or validation standard. For primary standards, pour necessary amounts of material into a sample container and run the sample per instrument manufacturer's user manual. Dispose of used nonhazardous materials. Do not pour back into bottle. These materials are not for dilution and do not require inversion or mixing to use unless otherwise stated on the bottle. Do not shake the standards as this can introduce greater error in the results. For secondary standards, place the vial in the instrument and run the sample per the manufacturer's user manual.

Not for direct use in food, cosmetics, finished pharmaceuticals or drug products. Supplier is not responsible for compliance with FDA Current Good Manufacturing Practices (cGMP), including without limitation for those finished drug products in 21 CFR Parts 210 and 211. Consult warranty limitations at www.gfschemicals.com

24029602