



North Central Laboratories

P.O. Box 8, Birnamwood, WI 54414

Certificate of Analysis

Product Number: YSI-3843
Product Name: 100 mg/L NH₄⁺-N STANDARD
Lot Number: 26A1F
Expiration Date: 07/02/27

The above material is certified to have a concentration of 100 ± 1.0 mg/L as N.

Mark Mahoney
Mark Mahoney
Chemist
NCL of Wisconsin Inc.

01/09/26
Date

The above material has been tested and meets the stated specifications after manufacture. Conditions of storage and use are beyond our knowledge and control. The possibility of contamination by the end user exists. Stated expiration date is based on our experience with retained samples, and is not to be construed as a warranty.

1.800.648.7836
Fax: (715) 449.2454
www.nclabs.com
Email: nclabs@nclabs.com