

# Certificate of Analysis



**YSI Incorporated**

Yellow Springs Instrument Co., Yellow Springs, Ohio 45387 USA

Phone 937 767-7241 • 800 765-4974 • Fax 937 767-9353

Product Category            Membrane  
Product Model Number    2754  
Lot Number                 17K100568  
Method                      Amperometric testing with standards

Time of Testing	Determination	Specification	Results
Initial	Average Calibration Sensitivity	8.0-40.0 nA	10.0 nA
	Average Linearity	9.50-10.50mmol/L	9.85 mmol/L
	Potassium Ferrocyanide	< 0.4mmol/L	0.05 mmol/L

Conforms to acceptance criteria

Completed by:                     Amanda Mills 10/26/2017                    

Your Name



A handwritten signature in blue ink that reads 'Lisa M. Abel'.

Lisa M. Abel  
Director, Quality and Corporate  
Responsibility