

Certificate of Analysis



YSI Incorporated

Yellow Springs Instrument Co., Yellow Springs, Ohio 45387 USA

Phone 937 767-7241 • 800 765-4974 • Fax 937 767-9353

Product Category Membrane
Product Model Number 2329
Lot Number 18H100207
Method Amperometric testing with standards

Time of Testing	Determination	Specification	Results
Initial	Average Calibration Sensitivity	6.0-30.0 nA	13.3 nA
	Average Linearity	14.3-15.7mmol/L	15.00 mmol/L
	Potassium Ferrocyanide	-0.3 - 0.3mmol/L	0.03 mmol/L

Conforms to acceptance criteria

Completed by: Christy Woodford 8/15/2018
Your Name

The procedures used to manufacture and test this product have been verified by QA for GMP compliance.



Lisa M. Abel
Director, Quality and Corporate
Responsibility