

Certificate of Analysis



YSI Incorporated

Yellow Springs Instrument Co., Yellow Springs, Ohio 45387 USA

Phone 937 767-7241 • 800 765-4974 • Fax 937 767-9353

Product Category Bioanalytical Standard
Product Model Number 2715
Lot Number 19A100249
Method Comparative enzymatic testing of Production standard vs Metrology standard

| Chemistry Tested | Determination | Tolerance | Result |
|------------------|---|---------------|---------|
| GLUCOSE | (Average of Production)-(Average of Metrology) | ± 20 mg/L | 10 mg/L |
| | Coefficient of variation for production standard (CV,%) | $\leq 2.0\%$ | 0.4 % |

Conforms to acceptance criteria

Completed by: Kristin Stover

Date: 1/9/2019



A handwritten signature in blue ink, appearing to read 'Lisa M. Abel'.

Lisa M. Abel
Director, Quality and Corporate
Responsibility