

Certificate of Analysis



YSI Incorporated

Yellow Springs Instrument Co., Yellow Springs, Ohio 45387 USA

Phone 937 767-7241 • 800 765-4974 • Fax 937 767-9353

Product Category Membrane
Product Model Number 2754
Lot Number 19B100766
Method Amperometric testing with standards

| Time of Testing | Determination | Specification | Results |
|-----------------|---------------------------------|------------------|-------------|
| Initial | Average Calibration Sensitivity | 8.0-40.0 nA | 14.5 nA |
| | Average Linearity | 9.50-10.50mmol/L | 9.97 mmol/L |
| | Potassium Ferrocyanide | < 0.4mmol/L | 0.01 mmol/L |

Conforms to acceptance criteria

Completed by: Christy Woodford

Your Name



A handwritten signature in blue ink that reads 'Lisa M. Abel'.

Lisa M. Abel
Director, Quality and Corporate
Responsibility