

Certificate of Analysis



YSI Incorporated

Yellow Springs Instrument Co., Yellow Springs, Ohio 45387 USA

Phone 937 767-7241 • 800 765-4974 • Fax 937 767-9353

Product Category Membrane
Product Model Number 2754
Lot Number 19C100660
Method Amperometric testing with standards

Time of Testing	Determination	Specification	Results
Initial	Average Calibration Sensitivity	8.0-40.0 nA	11.2 nA
	Average Linearity	9.50-10.50mmol/L	9.99 mmol/L
	Potassium Ferrocyanide	< 0.4mmol/L	0.00 mmol/L

Conforms to acceptance criteria

Completed by: Sandy Crawford
 Your Name

3/27/2019



A handwritten signature in blue ink that reads 'Lisa M. Abel'.

Lisa M. Abel
Director, Quality and Corporate
Responsibility