

Certificate of Analysis



YSI Incorporated
Yellow Springs Instrument Co., Yellow Springs, Ohio 45387 USA
Phone 937 767-7241 • 800 765-4974 • Fax 937 767-9353

a xylem brand

Product Category Bioanalytical Standard
Product Model Number 2368 Glucose Std
Lot Number 19D100120
Method Comparative enzymatic testing of Production standard vs. Metrology standard

| Chemistry Tested | Determination | Tolerance | Result |
|------------------|--|-------------------|------------|
| Glucose | (Average of production) - (average of metrology) | ± 0.13 mmol/L | 0.05mmol/L |
| Glucose | Coefficient of variation for production standard (CV, %) | $\leq 2\%$ | 1.1% |

Conforms to acceptance criteria

Completed by: Heather Brooks

Date: 4/17/19

The procedures used to manufacture and test this product have been verified by QA for GMP compliance.



A handwritten signature in blue ink that reads 'Gregory W. Popp'.

Gregory W. Popp
Quality Manager, Analytics