

# Certificate of Analysis



**YSI Incorporated**

Yellow Springs Instrument Co., Yellow Springs, Ohio 45387 USA

Phone 937 767-7241 • 800 765-4974 • Fax 937 767-9353

Product Category            Membrane  
Product Model Number    2903  
Lot Number                 20D100625  
Method                      Amperometric testing with standards

Time of Testing	Determination	Specification	Results
Initial	Average Calibration Sensitivity	5-40 nA	23.4 nA
	Average Exclusion	0.00±0.5g/L	0.05 g/L

Conforms to acceptance criteria

Completed by: Christy Woodford  
Your Name

4/30/2020



A handwritten signature in blue ink that reads 'Gregory W. Popp'.

Gregory W. Popp  
Quality Manager Analytics